



May be used with:

- *UPC5100, all ranges
- *UPC5110, all ranges
- UPC5000, all ranges up to 100 PSI
- UPC5010, all ranges up to 100 PSI

Part #58492

***Note:** Prior to serial number A1300,
approximate date prior to 7-9-1990

- Unit connects between calibrator test port and test hose.
- Maximum working pressure is 100 psi, and maximum working temperature is 140° F.
- Always vent system pressure prior to connecting or disconnecting.
- Always disconnect from calibrator and test hose prior to cleaning.
- Virtually no maintenance required. Approximately every two months apply a light coating of fluorinated krytox grease (part #55593) to the quick-disconnect fitting (mates with test hose) inner seal.

Note: The easiest method is to apply krytox grease to the outside surface between sealing lip and end of the mating test hose fitting (do not get grease in nitrogen flow opening). Vent unit line pressure to atmosphere. Plug test hose fitting into pressure trap quick-connect fitting and twist fitting back and forth to transfer krytox grease.



Toll Free: 888-295-8475 • International: 715-234-9171 • FAX: 715-234-6967

WEB SITE: www.4condec.com